

Bionic Products Proprietary Ltd
 PO Box 555
 Robina
 Queensland 4226
 Australia.

Page 1 of 1

Certificate of Test

Electrical Safety for Medical Ioniser

Date of Issue: 6th January 2015
SGS Reference: ELS184033/2/C/MG/13
Details of Sample: Models: Elanra MkII
 Ratings: mains adapter input 100-240V 50/60Hz 0.4A, output 12V 1.0A
 dc. Ambient 10 to 40 °C.
 S/No. 1809-079
Date of Receipt: 20th November 2013
Date of Test: 15th January 2014 to 24th December 2014
Test Specification: EN 60601-1: 2006 +A11:2011
 Medical electrical equipment – General requirements for basic
 safety and essential performance
 Exclusions : Clause 11.7 Biocompatibility, Clause 14 PEMs, Clause 12.2 Usability, Clause 12.3
 Alarms, Clause 17 EMC and ISO 14971 Risk management process.
Issue: 1
Conclusion: The sample tested was found to comply with the Test Specification.



Test Engineer: M.W. Garfitt



Counter Signatory: Z. Ellahi



All enquires relating to this certificate must be directed to the Test Engineer

This document is issued by the Company subject to its General Conditions of Service accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.
 Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only.
 Tests marked * are not UKAS accredited.