

Bionic Products Proprietary Ltd
 Unit 11/15 John Duncan Court
 Varsity Lakes
 QLD 4227
 Australia.

Certificate of Test

Electrical Safety for Portable Air Ioniser

Date of Issue: 8th March 2016
SGS Reference: ELS205939/2/R/AD/15
Details of Sample: Models: Elanra MkIII, Part n. C1000083
 Ratings:
 Mains Adapter, IN: 100-240V~ 50/60Hz 0.4A - Out: 12VDC 1.0A - Class II
 or External DC power supply: 12V
 or Internally powered by 6LR61 battery 9V
 S/No. EA151000012

Date of Receipt: 22nd June 2015, 5th November 2015

Date of Test: 24th June 2015 to 8th March 2016

Test Specification: IEC 60601-1: 2005 + *AM1 (2012)
 Medical electrical equipment – General requirements for basic safety and essential performance
 Exclusions : Clause 11.7 Biocompatibility, Clause 12.2 Usability, and Clause 17 EMC

Issue: 1


Conclusion: The sample tested was found to comply with the Test Specification.
 (Note that there may be specific hazards with the accessible high voltage output and sharp needle points. It is the client's responsibility to address these.)

A. Di Noia

M.W. Garfitt

Test Engineer: A. Di Noia

Authorised Signatory: M.W. Garfitt



All enquires relating to this certificate must be directed to the Test Engineer

This document is issued by the Company subject to its General Conditions of Service accessible at <http://www.sgs.com/en/Terms-and-Conditions.aspx> and, for electronic format documents, subject to Terms and Conditions for Electronic Documents at www.sgs.com/terms_e-document.htm. Attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein. Any holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents. This document cannot be reproduced except in full, without prior written approval of the Company. Any unauthorized alteration, forgery or falsification of the content or appearance of this document is unlawful and offenders may be prosecuted to the fullest extent of the law.
 Unless otherwise stated the results shown in this test report refer only to the sample(s) tested and such sample(s) are retained for 28 days only.
 Tests marked * are not UKAS accredited.