

The following sample(s) was/were submitted and identified on behalf of the applicant as:

Applicant: Elanra Pty. Ltd, Bionic Products Pty. Ltd.
11/15 John Duncan Court, Varsity Lakes, Qld. Australia 4227

Manufacturer: Bionic Products Pty. Ltd.
11/15 John Duncan Court, Varsity Lakes, Qld. Australia 4227

Factory Arrow Australia Pty. Ltd.
46 Lathams Rd, Carrum Downs, Vic, 3201, Australia

Testing Laboratory: SGS-CSTC Standards Technical Services Co., Ltd. Guangzhou Branch
198 Kezhu Road, Science City, Economic & Technology Development Area,
Guangzhou, Guangdong, China

The standards: Clause 8.9.3.2 of EN 60601-1:2006

Test item description: Elanra Air Pendant Ioniser PCBA

Model/Type reference: Elanra Air Pendant Ioniser PCBA

Ratings: 5kV DC; 10uA

Test result: In the opinion of SGS – CSTC the presented appliance was found to be in compliance with the test specification as indicated in the details on the following pages.

Remark: None




Fly Lu
Project Reviewer


Aaron Cen
Project Engineer

Medical Electrical Equipment Laboratory

